

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, AS ON- 09.03.2019.

1. Name of Hospital : V.M.M.C. & SAFDARIJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 03
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non-functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	PMR	CAD- CAM	01 15.09.2008	July 2009	Rs. 62,00,000/-	Feb, 2016	Under CMC	Loss of USB Key	USB Key Approximate expenditure: ₹4,78,000/-	The firm has submitted the revised estimate of Rs. 4,78,000/- dated 24.12.2018 sent to HOD-PMR to verify and for needful on 26.12.2018. As desired by user department, justification for increase in price has been provided by firm, file send to HOD PMR to examine the justification on 23.01.19. File sent to HOD (PMR) to fill the Repair performa on 13.03.2019. The file is to be put up to Repair Committee.
2	PMR	Ultrasound Therapy Model: Medison	01 01.10.2010	Jan, 2011	Rs. 1,17,600/-	-----	No	Communication Cable is not working	Communication Cable is to be replaced. Approximate expenditure: ₹86,000/-	Query raised by inspection committee. Letter sent to the firm for clarification on 18.03.2019. The firm supplied the item in Medical Store on 08.02.2019. File send to HOD-PMR to fix the date of inspection on 08.02.19.
3.	Microbiology	Horizontal Steam Sterilizer-YSU 608	01 25.10.2011	03.01.2012	₹17,12,550/-	October, 2018	No	Technical	Approximate expenditure: ₹65,000/-	Inspection under process. The Firm M/s Yorco Sales Pvt. Ltd. has submitted estimate for repair, the estimate sent to HOD Microbiology for verification and to fill the form of repair on 21.12.2018

