

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, for the month January, 2019.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)  
 3. Total Number of Non-Functional Equipments : 03  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	PMR	Shoulder Elbow CPM	01 31.03.2011	20.05.2011	USD 7500	21.03.2014	No	Repair required	To be obtained from the firm	Installed and functional.
2	PMR	CAD- CAM	01 15.09.2008	July 2009	Rs. 62,00,000/-	Feb., 2016	Under CMC	Loss of USB Key	USB Key	As informed by the firm (M/s Re Genesis) the new version of Omega cannot be run with configuration in existing computer, the cost involved the computer would be extra and was not the part of previous quotation. Therefore, firm has submitted the revised estimate of Rs. 4,78,000/- dated 24.12.2018 sent to HOD-PMR to verify and for needful on 26.12.2018
3	PMR	Ultrasound Therapy Model: Medison	01 01.10.2010	Jan, 2011	Rs. 1,17,600/-	-----	No	Communication Cable is not working	Communication Cable is to be replaced.	Supply order issued to firm dated 10.11.2018 with a delivery period up to 15.12.2018. The firm submitted letter dated 08.01.2019 for extension till 15.02.2019 due to some unforeseen reasons.
4.	Microbiology	Horizontal Steam Sterilizer-YSU 608	01 25.10.2011	03.01.2012	₹17,12,550/-	October, 2018	No	Technical	-	The Firm M/s Yorco Sales Pvt. Ltd. has submitted estimate for repair, the estimate sent to HOD Microbiology for verification and to fill the form of repair on 21.12.2018