

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, for the month June, 2018.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 05
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	PMR	Shoulder Elbow CPM	01 31.03.2011	20.05.2011	USD 7500	21.03.2014	No	Repair required	To be obtained from the firm	Estimate of ₹2,80,000/- submitted by the Firm and put up before Repair Committee held on 04.06.2018, Repair Committee directed to sent a letter to firm to appear before repair committee for negotiation as the repair cost is more than 50 % of the equipment cost. Letter sent to the firm on 05.06.18
2	PMR	Comprehensive Computerized Hand & Upper Extremety Evaluation System	01 08.04.2009	08.07.2010	Rs. 8,06,768/-	2011	NO	Repair required	To be obtained from the firm	Letter Sent to the firm on 03.04.2018 to submit the estimate. Firm's engineer has inspected the machine, submitted the service report near about 07.05.2018 and informed that they will submit the estimate in 2-3 days. However, the firm has not yet submitted the estimate and a reminder letter has been sent to the firm on 05.06.2018 to submit the estimate and authorization

										certificate.
3	PMR	CAD- CAM	01 15.09.2008	July 2009	Rs. 62,00,000/-	Feb., 2016	Under CMC	Loss of USB Key	USB Key	As requested by the HOD (PMR), a letter sent to the firm (M/s RE Genesis Healthcare Service) on 05.06.2018, to inspect the machine and submit the final estimate.
4	PMR	Ultrasound Therapy Model: Medison	01 01.10.2010	Jan, 2011	Rs. 1,17,600/-	-----	No	Communication Cable is not working	Communication Cable is to be replaced	Estimate of ₹89,900/- submitted by the Firm and put up before Repair Committee held on 04.06.2018, Repair Committee directed to sent a letter to firm to appear before repair committee for negotiation as the repair cost is more than 50 % of the equipment cost. Letter sent to the firm on 05.06.18
5	Surgery	DVT Pump Model: DVT 2600	01 20.03.2014	20.03.2014	Rs. 1,25,475/-	27.03.2018	Under AMC	Short Circuit Fused not working	Short Circuit Fused is to be replaced	Supply order for repair of the equipment has been placed to the firm on 12.06.2018.

C.M.O. (Medical Store)