

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON- 12.09.17.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 04 + 02(to be condemned)
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non-functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	The firm has given a letter that the equipment has to be sent to the Japan for repair purpose but the monitor is very old and due to its import and export documentation issues they are unable to sent it to the Japan. However the firm has submitted a proposal of giving a working monitor of same model as a replacement for the defective monitor against the same service order. A letter to firm has been sent on 11.09.2017 to clarify whether the defective monitor will be returned to hospital or is it a buy back proposal and whether any payment will required to be made by the hospital.
2	Medicine	Ventilator M/s. Instromedix	01	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Three Ventilators repaired and installed on 06.09.2017. One Ventilator likely to be repaired within 4-5 days.
3	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	As per direction of Dte.G.H.S. HOD (Dental) has been issued an office note

		New firm M/s. Hosta care								to return the file, copy to the Medical Superintendent of this hospital with a request to intervene in the matter on 02.08.2017. The file has been received from the user department, wherein HOD (Dental has informed that the repair cost of the equipment (Rs. 1,31,900/-) is more than the purchase cost of the equipment (Rs. 1,28,960/-). File has been put up for further orders/directions. In view of the long pending non-functional status of the equipment and higher repair cost, a proposal to condemnation of the equipment will be forwarded to M.S. by M&E Section.
4	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	To be Condemned.
5	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	To be Condemned.

C.M.O. (Medical Store)