

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTINS MORE THAN RS. ONE LAKH, FOR THE MONTH OF – March, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 1569
 3. Total Number of Non-Functional equipment : 01
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioningof equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Cardiology	Cath Lab	One	2002	Euro 696800	26.02.14	Yes, CAMC	Spare required	Cooling system broken down	Functional
2.	Radiology	100MA portable X-ray machine	One	1998	1.30 lacs	Feb,14	Yes	Spare required	Hand Switch	Functional
3.	Pediatrics Ward-18	Bear Cub Ventilator	One	2009	USD 14520	March,14	No	Spares required	Flow Sensor	Estimate has been approved by Repair committee and file has been sent to Spl. D.G. & M.S. on 15.4.14 for expenditure sanction.

Sd/-
C.M.O. (Medical Store)