

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF –June, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
2. Total Number of Equipment : 2693 (updated up to 31.03.14)
3. Total Number of Non-Functional equipment : 05
4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non- functioningof equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1.	Eye	Scrub Station M/s. Rudex International	02	2010	1,90,000/-	April, 14	No	Spares required	Sencer, foot pump, etc.	Supply order placed on – 09.07.14
2.	Anesthesia	Avea Ventilator M/s. Rohanika Electronics	03	2011	14.00 lac	April,14	yes	Battery Assy. Reqd.	Battery Assy.	Supply order placed on- 12.06.14, reminder given to the firm on 15.07.14

Sd/-
C.M.O. (Medical Store)

