

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-09.05.16.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3441 (Updated up to 31.03.16)
 3. Total Number of Non-Functional Equipments : 16
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	One	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	Letter dated-29.03.16 received from the principal firm M/s. Italray. File has been sent to HOD (Radiology) for comments.
2	Radiotherapy	HDR Brachytherapy machine	One			December,2009	No	Technical restrictions from AERB		File has been put up to M.S. for Administrative approval & Expenditure sanction.
3	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Fresh Estimate received from the firm and file sent to HOD (CTVS) for verification of estimate and filling up of requisition form.
4	Resp. Medicine	Video Thorow scope M/s. Olympus Medical	01	26.06.12	33.00 lacs	Nov, 2015	Yes	Spares required	-----	S. Order placed on 09.02.16. Firm applied for extension upto 15.05.16.
5	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	Dec.2015	Under warranty	Compressor not working	-----	Firm has intimated that they have placed the order for spare parts to the principal firm and will repair the machine shortly. Reminder sent to the firm on 8.04.16.
6	Medicine	Ventilator	02	August,2009	4.00 lac	January,2016	No	Spares required	Spares required	Estimate received from the

		M/s. Instromedix			each						firm and sent to HOD (Med.) last week for verification and filling up of requisition form.
7	Dental	Sand Blaster M/s. IDS Denmed	01	23.02.10	2.50 lacs	2011	No	Firm has not provided the service during warranty period	Spares required		Estimate has been received from the firm and matter for repair is under process.
8	Dental	Ceramic Furnace M/s. International Dental System	01	1.07.10	5.04 lacs	2011	No	- do -	- do -		- do -
9	Dental	Duplicating Unit M/s. IDS Denmed	01	01.07.10	2.01 lacs	2011	No	- do -	- do -		- do -
10	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -		- do -
11	Pediatrics Ward-18	Transport Ventilator Elisee-150 M/s. Instromedix	01	2009	3.95 lacs	December,2015	No.	Spares required	Turbine, Rotary valve		Supply order placed to the firm on 09.05.16
12	Pediatrics Ward-20	Transport Ventilator M/s. Instromedix	01	2009	3.95 lac	January,16	No	Spares required	Turbine, Rotary valve, Peep motor, oxygen sensor		Firm has submitted the required documents, and file has been put up to HOD (Peads.) for verification of estimate and filling up of requisition form.
13	Microbiology	CO2 Incubator	01	2014	4.93 lac	April,16	Yes	-----	-----		Letter sent to firm on 19.04.16 and reminder on 09.05.16
14	Microbiology	Vertical Autoclave	01	2013	20.00 lac	April,16	Yes	-----	-----		Letter sent to firm on 09.05.16

SD/-
C.M.O. (Medical Store)