

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF FEBRUARY-2016

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3355 (Updated up to 31.12.15)
 3. Total Number of Non-Functional Equipments : 07
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	One	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	Letter sent to RML Hospital to obtain source of repair (if any) and file sent to HOD (Radiology) on 04.03.16 for suggesting further line of action and other source of repair.
2	Radiotherapy	HDR Brachytherapy machine	One			December,2009	No	Technical restrictions from AERB		Meeting of the committee held on 16.02.16. Letter sent to firm on 17.02.16 for submitting repair estimate.
3	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Estimate received from the firm and file sent to HOD CTVS on 02.02.16.
4	Resp. Medicine	Video Thorow scope M/s. Olympus Medical	01	26.06.12	33.00 lacs	Nov, 2015	Yes	Spares required	-----	S. Order placed on 09.02.16
5	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	Dec.2015	Under warranty	Compressor not working	-----	Firm has intimated that they have placed the order for spare parts to the principal firm and will repair the machine shortly.

6	Medicine	Ventilator M/s. Instromedix	02	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Approval of M.S. has been obtained to send the machine to the company workshop. File has been sent to the HOD (Medicine) for further needful.
---	----------	--------------------------------	----	-------------	---------------	--------------	----	--------------------	--------------------	--

Sd/-
C.M.O. (Medical Store)

