

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF –July, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 2693 (updated up to 31.03.14)
 3. Total Number of Non-Functional equipment : 13
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non-functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1.	Eye	Scrub Station M/s. Rudex International	02	2010	1,90,000/-	April, 14	No	Spares required	Sencer, foot pump, etc.	Supply order placed on – 09.07.14, Reminder sent to firm on 12.8.14
2.	Anesthesia	Avea Ventilator M/s. Rohanika Electronics	03	2011	14.00 lac	April,14	yes	Battery Assy. Reqd.	Battery Assy.	Functional
3.	Radiotherapy	Cobalt 858 M/s. Kirloskar	One	2000	2.00 crore	Partially functional	Yes	-----	-----	Letter sent to firm on 6.8.14, regarding deputing Sr. Technician from company.
4	Medicine	Defibrillator M/s. A.Square	2	2006	4.00 lacs	June,14	Yes	Spares required	Paddles	Supply order placed on 12.8.14
5	C.I.O	C-Arm M/s. Wipro GE	One	2013	Rs.24.00 lacs	June,14	Under warranty	Part to be changed	Assembly, CCD Camera with CCU, EVR	Letter & fax sent to the firm on- 5.8.14. File has been sent to M.S. for further direction against company as they have given an estimate inspite of equipment being under warranty.
6	CTVS	Sternal Saw M/s. J.Mitra & Bros.	One	2008	Rs.4.19 lacs	June,14	AMC	Spare part required	Motor, Inner core	Approved by the repair committee, order to be placed
7	CTVS	Ventilator M/s. Rohanika	02	2010	Rs.19.72 lacs	June,14	AMC	Spare part required	Main PCB	Approved by repair committee, order to be

		Electronics								placed.
8	Resp. Medicine	Defibrillator M/s. Biomed Instruments	02	2012	9.00 lacs	June,14		Accessories required	Battery	Approved by the repair committee, order to be placed.
9	Cardiology	ECG Machine M/s. A.Square Systems	01	2005	99000/-	June,14		Spares required	Battery, LCD, paper loading unit	Approved by repair committee, order to be placed.
10	Ward-A	Portable ventilator M/s. Rudex International	01	2011	3.9 lacs	July,14	No	Spares required	Main power supply system	Supply order placed on 11.08.14

C.M.O. (Medical Store)

