

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF –October, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 2824 (updated up to 30.09.14)
 3. Total Number of Non-Functional equipment : 08
 4. Theft or Fire Occurred : Nil

| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 | 11 |
|-------|----------------|--|---|-----------------------------------|-------------------------|--|--|--|---|---|
| S.No. | Name of Deptt. | Name of equipment Department wise | Total No. of equipment & date of purchase | Date of installation of equipment | Cost of equipment (Rs.) | Date from which equipment(s) is/are non - functional | Whether the equipment(s) is under AMC/warranty, if not reasons thereof | Reasons for non-functioning of equipment | Details of part required for repair along with estimate expenditure | Action taken to repair the equipment & approximate time by which equipment will become functional |
| 1. | Eye | Scrub Station M/s. Rudex International | 02 | 2010 | 1,90,000/- | April, 14 | No | Spares required | Sencer, foot pump, etc. | Functional |
| 2 | C.I.O | C-Arm M/s. Wipro GE | One | 2013 | Rs.24.00 lacs | June,14 | Under warranty | Part to be changed | Assembly, CCD Camera with CCU, EVR | Spares received, under installation. |
| 3 | Resp. Medicine | Defibrillator M/s. Biomed Instruments | 02 | 2012 | 9.00 lacs | June,14 | No | Accessories required | Battery | Supply order placed on-1.09.14 Reminder sent on 10.10.14 Extension given upto-17.11.14. |
| 4 | SIC | Battery Power system | 2 | 2011 | ----- | August,14 | AMC | Spares required | Spares required | Spares received, under installation. |
| 5 | SIC | Arthroscopy shaver system | 3 | 2011 | 34.61 lacs | August,14 | AMC | Spares required | Spares required | Spares received, under installation. |

Sd/-
C.M.O. (Medical Store)