

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF –December, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 2900 (updated up to 31.12.14)
 3. Total Number of Non-Functional equipment : 02
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non-functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Resp. Medicine	Defibrillator M/s. Biomed Instruments	2	2012	9.00 lacs	June,14	No	Accessories required	Battery	Functional
2	SIC	Battery Power system M/s. Cure Surgical	2	2011	-----	August,14	AMC	Spares required	Spares required	Functional
3	SIC	Arthroscopy shaver system M/s. Cure Surgical	3	2011	34.61 lacs	August,14	AMC	Spares required	Spares required	Functional
4	ENT	O.T.Table M/s. Hospimedica International	1	2001	15.50 lacs	November,14	AMC	Spares required	Spares required	Supply order placed on 22.12.14.
5	Radiology	1000 MA Digital X-Ray Machine M/s .Empire Instrumentation	1	2007	Euro 296000	Frequently gone out order and now continuously non functional.	Under Warranty	Some parts to be changed	Spares required	Earlier matter was taken up by the HOD directly with the firm, now letter has also been sent to the firm on 4.12.14 to make the machine functional, immediately. Reminder sent on 19.12.14 and 12.01.14

(Dr. R.P.Arora)
 C.M.O. (Medical Store)