

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-11.04.17.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3622 (Updated up to 31.12.16)  
 3. Total Number of Non-Functional Equipments : 09  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	An inspection was done by Dte.G.H.S. representative (Dr. Anil Sain, Addl .Dy. Dir. General) on 22.02.2017 and advised to initiate the process of blacklist the foreign firm and also advised the department of radiology to look into the possibilities of getting the equipment condemned as the equipment is 10 years old and it is difficult to get the equipment repaired from the foreign firm. Copy of the inspection report has been received on 10.03.17, The file is lying with the user department (Radiology). It is informed by the user deptt. on 10.04.2017 that they have started the condemnation process, and will provide the files as and when their work with the file will finished. The process of blacklisting of the firm will be initiated as and when the file will be returned by the user department.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	----	Firm has re-inspected the equipment and submitted a revised estimate of repair with request for permission to collect the equipment to send it to their Japan Tech. services. Repair Committee has approved the proposal.

										The firm has been requested vide letters Dt. 14.02.17, 03.03.17, emailed dt. 03.03.2017 and letter dated 10.04.2017 to deposit collateral security towards cost of equipment.
3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Repair order placed on-15.12.16. Firm has asked and given extension till-15.03.17. It is telephonically informed by the firm that they will provide the required spares by the 17.04.2017.
4	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	M/s. IDS Denmed Pvt. Ltd. has confirmed vide letter dated 28.03.2017 that they are no more authorized by the manufacturer to repair the equipment and M/s. Dentsplay India is now authorized. A letter has been issued to HOD (Dental) to inform the address of M/s Dentsplay India and also to inform the efforts made by him to make the equipment functional.
5	Blood Bank	Automatic cell counter M/s. Beckman coulter	01	2014	10.39 lac	Nov, 16	Under warranty till 19.12.16	-----	-----	Letter sent to firm on 07.12.16 for repair. Reminder sent to the firm on 03.01.17 and 02/17. It is telephonically informed by the user depts. that the firm had repaired the equipment but, it has again become non-functional. Service Engineers are trying to repair the equipment.

**C.M.O. (Medical Store)**