

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, for the month April 2018.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)  
 3. Total Number of Non-Functional Equipments : 05  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	PMR	Shoulder Elbow CPM	01 31.03.2011	20.05.2011	USD 7500	21.03.2014	No	Repair required	To be obtained from the firm	Letter Sent to the firm on 03.04.2018 to submit the estimate
2	PMR	Comprehensive Computerized Hand & Upper Extremety Evaluation System	01 08.04.2009	08.07.2010	Rs. 8,06,768/-	2011	NO	Repair required	To be obtained from the firm	Letter Sent to the firm on 03.04.2018 to submit the estimate
3	PMR	CAD- CAM	01 15.09.2008	July 2009	Rs. 62,00,000/-	Feb., 2016	Under CMC	Loss of USB Key	USB Key	CAD-CAM Machine is under custody of Sh. Sohan Pal, Workshop Manager. Machine is non functional since Feb.,2016 due to loss of USB Key. In this regard, enquiry has been done against Sh. Sohan Pal and he was exonerated from the charges leveled against him vide order no. 6-12/2016-Vig. dated 14.10.2017. Firm M/s. Hospimedica international submitted the repair /Software upgradation estimate of Rs. 6,93,000/- dated 19.08.2016. Moreover, the same firm revised its quotation and enhanced

										the estimate of Rs. 8,72,000/-, dated 08.02.2018. However, another firm M/s. RE Genesis Healthcare Services submitted the quotation No. REG/2017-18/00160 dated 12.09.2017 of Rs. 97,600/- for the same. Same is under Consideration.
4	PMR	Ultrasound Therapy Model: Medison	01 01.10.2010	Jan, 2011	Rs. 1,17,600/-	-----	No	Communication Cable is not working	Communication Cable is to be replaced	Letter Sent to the firm on 09.04.2018 to submit the estimate
5	Surgery	DVT Pump Model: DVT 2600	01 20.03.2014	20.03.2014	Rs. 1,25,475/-	-----	Under AMC	Short Circuit Fused not working	Short Circuit Fused is to be replaced	File put up to the user department on 05.04.2018 for fill up the requisite form for repair.

**C.M.O. (Medical Store)**