

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON- 08 .08.17.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 09
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	The user department has sent their recommendations to the chairman, condemnation committee for condemnation of the equipment on 29.05.2017. Next condemnation committee has been supposed to be held in mid of July, 2017 but could not be held. The condemnation of process will be done as soon as condemnation committee will be held. A letter to Dte.G.H.S. for blacklisting of the foreign principal has been sent on 03.07.2017.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	----	The firm has given a letter that the equipment has to be sent to the Japan for repair purpose but the monitor is very old and due to its import and export documentation issues they are unable to sent it to the Japan. However the firm has submitted a proposal of giving a working monitor of same model as a replacement for the defective monitor against the same service order. The file alongwith proposal has

										been sent to the HOD (CTVS) for comments on 02.08.2017, but no reply received till date.
3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	The firm has informed that equipment cannot be repaired in the hospital and they have to take the equipment to their workshop. Accordingly user department has issued letter o the firm to collect the equipment o 04.08.2017
4	Dental	Apex Locator M/s. IDS Denmed New firm M/s. Hosta care	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	As per direction of Dte.G.H.S. HOD (Dental) has been issued an office note to return the file, copy to the Medical Superintendent of this hospital with a request to intervene it the matter on 02.08.2017. The file has been received from the user department, wherein HOD (Dental has informed that the repair cost of the equipment (Rs. 1,31,900/-) is more than the purchase cost of the equipment (Rs. 1,28,960/-). File has been put up for further orders/directions.
5	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	The user department has sent dully filled proforma for condemnation. The condemnation of process will be done as soon as condemnation committee will be held. The process of blacklisting of the firm is under process.

C.M.O. (Medical Store)