

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, for the February, 2018.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 02 + 01(to be condemned)
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	The firm has informed, vide letter dated 21.11.2017, that they had dispatched the defective monitor to Japan for repair as they had found its import and export documentation which earlier they were not able to trace it. They will return the monitor back after completion of repair. However they have sought extension of three months as it will take time in transition between India & Japan. The extension till 20.02.2018 has been granted vide letter dated 30.11.2017.
2	Dental	Apex Locator M/s. IDS Denmed New firm M/s. Hosta care	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	The equipment has been repaired and working satisfactory.

3	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	To be Condemned.
4	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	The equipment has been purposed by user department to be condemned; However, as per fresh direction received from the Medical Superintendent, an open tender for the repair of the equipment has been published on 04.01.18 due on 19.01.2018, to get the possibility of repair the equipment. Due date extended up to 30.01.2018 as no tender was received, no tender received on 30.01.2018 as well, hence tender has been closed and file has been sent to HOD to initiate condemnation process.

C.M.O. (Medical Store)