

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-13.06.17.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)  
 3. Total Number of Non-Functional Equipments : 09  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non-functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	The user department has sent their recommendations to the chairman, condemnation committee for condemnation of the equipment on 29.05.2017. Next condemnation committee is supposed to be held in mid of July, 2017. A letter to Dte.G.H.S. for blacklisting of the foreign principal has also prepared and submitted for approval of Medical Superintendent which may be dispatch soon.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Supply Order to the firm to repair the equipment issued on 01.06.2017 with the condition to provide the backup equipment. The time given to the firm to supply these spares is 12.07.2017. The firm has collected the defective equipment for repair at their service centre and provided a backup equipment on 12.06.2017

3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Spares for the repair of ventilators (04 Nos.) have been received in the store on 15.05.17. However, inspection of the spares is pending. A letter has been sent to the HOD on 12.06.2017 to coordinate with the firm, for an urgent inspection in order to get the equipment repaired.
4	Dental	Apex Locator M/s. IDS Denmed  <b>New firm</b> <b>M/s. Hosta care</b>	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	M/s. Hosta care has submitted the quotation for repair. The file has been sent to HOD Dental on 25.04.17 to examine the estimate. File is yet to be received back from HOD. Reminders have been sent to HOD on 17.05.17 and 09.06.17
5	Blood Bank	Automatic cell counter M/s. Beckman coulter	01	2014	10.39 lacs	Nov, 16	Under warranty till 19.12.16	-----	-----	Repair done. Equipment is functioning.
6	Pathology	Fully Automated Blood cell counter  M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	The file has been sent to user department on 24.04.17 to initiate and follow up the condemnation process of the equipment and process of blacklisting of the firm. User department has informed that the equipment is under process for condemnation. A letter has been sent to the HOD to get the present status on 12.06.2017

**C.M.O. (Medical Store)**