STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, AS ON- 19.03.2019.

Name of Hospital

Total Number of Equipments

Total Number of Non-Functional Equipments:

Theft or Fire Occurred

V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI

3849 (Updated up to 31.03.17)

<u>≅</u> 03

4.	ω	2	S.No.
Microbiology	PMR	PMR	Name of Deptt.
Horizontal Steam Sterilizer-YSU 608	Ultrasound Therapy Model: Medison	CAD- CAM	Name of equipment Department wise
01 25.10.2011	01 01.10.2010	01 15.09.2008	Total No. of equipment & date of purchase
03.01.2012	Jan, 2011	July 2009	Date of installation of equipment
₹17,12,550/-	Rs. 1,17,600/-	Rs. 62,00,000/-	6 Cost of equipment (Rs.)
October, 2018		Name of Name	
No	No	Under CMC	Whether the equipment(s) is under AMC/warranty, if not reasons
Technical	Communication Cable is not working	Loss of USB Key	9 Reasons for non functioning of equipment
Approximate expenditure: ₹65,000/-	Communication Cable is to be replaced. Approximate expenditure: ₹86,000/-	USB Key Approximate expenditure: ₹4,78,000/-	Details of part required for repair along with estimate expenditure
The Firm M/s Yorco Sales Pvt. Ltd. has submitted estimate for repair, the estimate sent to HOD Microbiology for verification and to fill the form of repair on 21.12.2018 Put in repair committee on 08.02.19, approved & put up to Medical Superintendent for expenditure approval on 08.02.19.	The firm supplied the item in Medical Store on 08.02.2019. File send to HOD-PMR to fix the date of inspection on 08.02.19. Query raised by inspection committee. Letter sent to the firm for clarification on 18.03.2019.	firm has submitted ed estimate of Rs. 4,78,0 d 24.12.2018 sent to H to verify and for needful 2.2018. It to verify and for needful 2.2018. It is performed by different estimates on 13.03.201.	Action taken to repair the equipment & approximate time by which equipment will become functional