

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-14.03.17.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3622 (Updated up to 31.12.16)  
 3. Total Number of Non-Functional Equipments : 09  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	Letter date-31.08.16 has been written to M/s. Italray by the Dy. Director (Proc.) to repair the machine on priority basis without any additional expenditure. Comments of HOD received, Reminder has been sent to M/s. Italray in reference to letter dt.31.08.16. Reminder sent to firm on-16.11.16. File sent to HOD (Radiology) for comments, whether firm has responded or not. It is informed by the HOD (Radiology) that firm M/s. Italray has not responded for repair of the equipment. An inspection was also done by Dte.G.H.S. representative (Dr. Anil Sain, Addl .Dy. Dir. General) on 22.02.2017 and advised to initiate the process of blacklist the foreign firm and also advised the department of radiology to look into the possibilities of getting the equipment condemned as the equipment is 10 years old and it is difficult to get the equipment repaired from the foreign firm. Copy of the inspection report has been received on 10.03.17, process will be initiated soon.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	----	Firm has re-inspected the equipment and submitted a revised estimate of repair with request for permission

										to collect the equipment to send it to their Japan Tech. services. Repair Committee has approved the proposal. The firm has been requested vide letters Dt. 14.02.17 & 03.03.17 and also emailed on 03.03.2017, to deposit collateral security towards cost of equipment.
3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Repair order placed on-15.12.16. Firm has asked and given extension till-15.03.17. This store has not received the spares till date.
4	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	Reminder sent to the firm on 10.12.16, 09.01.17, 07.02.17 and 04.03.17 to submit repair estimate with requisite papers. A Note to HOD Dental has also been sent on 10.02.2017 and 04.03.17 to inform the steps taken for repair of equipment. HOD Dental has telephonically informed on 14.03.17 that the M/s. IDS denmed pvt. Ltd. is no more authorized by the manufactured for repair of the equipment and therefore HOD dental is trying to take up the matter with authorized firm.
5	Blood Bank	Automatic cell counter M/s. Beckman coulter	01	2014	10.39 lac	Nov, 16	Under warranty till 19.12.16	-----	-----	Letter sent to firm on 07.12.16 for repair. Reminder sent to the firm on 03.01.17 and 02/17. Firm representative visited the department and equipment is under observation.

**C.M.O. (Medical Store)**