

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, for the month May & June, 2019.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)  
 3. Total Number of Non-Functional Equipments : 03  
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	PMR	CAD- CAM	01 15.09.2008	July 2009	Rs. 62,00,000/-	Feb., 2016	Under CMC	Loss of USB Key	USB Key  Approximate expenditure: ₹4,78,000/-	Supply order given to firm on 08.05.2019, Installation of Software is under process.
2	PMR	Ultrasound Therapy Model: Medison	01 01.10.2010	Jan, 2011	Rs. 1,17,600/-	-----	No	Communication Cable is not working	Communication Cable is to be replaced.  Approximate expenditure: ₹86,000/-	Inspection done, item indented by the user department on 02.05.19 and installation is under process.
3.	Microbiology	Horizontal Steam Sterilizer-YSU 608	01 25.10.2011	03.01.2012	₹17,12,550/-	October, 2018	No	Technical	-  Approximate expenditure: ₹65,000/-	Supply order issued on 19.02.2019, the firm supplied the items on 26.02.2019, Inspection is under process, the inspection committee has asked some clarifications from firm on 25.04.19 and asked some query from HOD Microbiology on 04.06.2019. The file was received from Inspection committee on 11.06.2019 and is being sent to HOD for answer of the queries asked by inspection committee.

**C.M.O. (Medical Store)**